

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002241  
12:22 11/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:51  
PRESSURIZING 1:58  
STERILIZING 3:08  
136.5 C 226kPa 3:08  
136.3 C 226kPa 3:39  
136.4 C 226kPa 4:09  
136.3 C 224kPa 4:39  
136.3 C 225kPa 5:09  
136.4 C 222kPa 5:39  
136.1 C 223kPa 6:09  
136.4 C 226kPa 6:39  
136.4 C 226kPa 6:39  
VENTING 6:39  
AIR DRYING 7:36  
CYCLE COMPLETE 37:36  
-----

Digital Signature #  
1285FA4A8D10F239  
-----