

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1568

Name: RITTER2  
Sterilizer ID: V2313228  
10/05/2022 02:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.6 F 0.0  
2:00 187.4 F 0.1  
4:00 193.7 F 0.7  
6:00 214.6 F 4.9  
8:00 225.4 F 9.7  
10:00 233.8 F 9.0  
12:00 240.7 F 12.4  
14:00 240.1 F 11.1  
16:00 239.3 F 10.4  
18:00 253.1 F 17.2  
20:00 264.8 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.8  
1:00 272.6 F 29.4  
2:00 271.7 F 28.7  
3:00 272.3 F 29.6  
4:00 272.3 F 29.1

Min 271.4 F 28.6  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.2 F 0.0  
5:00 230.3 F 0.0  
10:00 241.6 F 0.0  
15:00 234.4 F 0.0  
20:00 223.7 F 0.0  
25:00 223.5 F 0.0  
30:00 224.1 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 21:54  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:00:06  
CYCLE COMPLETE

10/05/2022 03:12 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_