
Midmark M11 - v1.0.5
Total Cycles: 3528

Name: RITTER2
Sterilizer ID: V2313228
04/15/2024 01:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.9 F 0.0
2:00 146.0 F 2.6
4:00 167.0 F 5.7
6:00 205.6 F 9.6
8:00 205.7 F 9.1
10:00 222.4 F 6.7
12:00 231.9 F 11.1
14:00 230.4 F 7.8
16:00 244.5 F 13.5
18:00 243.8 F 12.6
20:00 255.9 F 18.9
22:00 266.8 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.9 F 29.7
2:00 272.0 F 29.1
3:00 271.2 F 29.0
4:00 272.5 F 29.8

Min 271.2 F 28.7
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.1 F 0.0
5:00 197.2 F 0.0
10:00 201.0 F 0.0
15:00 199.3 F 0.0
20:00 196.7 F 0.0
25:00 193.8 F 0.0
30:00 190.3 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 23:35
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:02:14
CYCLE COMPLETE

04/15/2024 02:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____