
Midmark M11 - v1.0.0
Total Cycles: 2992

Name: RITTER 1
Sterilizer ID: V2178198
03/09/2023 04:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.8 F 0.0
2:00 193.0 F 4.3
4:00 209.7 F 9.6
6:00 221.1 F 5.9
8:00 232.7 F 10.5
10:00 232.9 F 7.3
12:00 242.8 F 11.4
14:00 246.4 F 13.4
16:00 259.4 F 20.5
18:00 270.8 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.4
1:00 272.7 F 29.0
2:00 271.6 F 28.3
3:00 272.6 F 28.9
4:00 271.7 F 28.3

Min 271.6 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 234.1 F 0.0
10:00 231.2 F 0.0
15:00 225.0 F 0.0
20:00 224.6 F 0.0
25:00 226.0 F 0.0
30:00 226.9 F 0.0
Drying Complete

FILLING: 2:15
HEATING: 18:36
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:05
CYCLE COMPLETE

03/09/2023 05:30 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____