
Midmark M11 - v1.0.0
Total Cycles: 4296

Name: RITTER 1
Sterilizer ID: V2178198
02/13/2024 08:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.5 F 0.0
2:00 82.2 F 0.2
4:00 110.8 F 1.1
6:00 113.2 F 3.4
8:00 137.5 F 6.0
10:00 168.5 F 9.1
12:00 187.7 F 12.9
14:00 209.1 F 3.6
16:00 218.0 F 6.8
18:00 227.4 F 10.4
20:00 232.1 F 8.3
22:00 237.5 F 10.3
24:00 235.2 F 8.3
26:00 247.0 F 13.7
28:00 258.2 F 19.7
30:00 268.3 F 26.0
32:00 271.5 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.8 F 29.0
2:00 271.7 F 28.7
3:00 272.7 F 28.9
4:00 271.7 F 28.8

Min 271.4 F 28.4
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.5 F 0.0
5:00 233.5 F 0.0
10:00 230.8 F 0.0
15:00 229.4 F 0.0
20:00 229.2 F 0.0
25:00 227.5 F 0.0
30:00 229.4 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 32:02
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:10:16
CYCLE COMPLETE

02/13/2024 09:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____