
Midmark M11 - v1.0.0
Total Cycles: 1994

Name: RITTER 1
Sterilizer ID: V2178198
06/06/2022 12:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.7 F 0.0
2:00 141.9 F 2.3
4:00 163.4 F 5.6
6:00 190.5 F 9.8
8:00 219.5 F 6.2
10:00 219.6 F 8.3
12:00 232.8 F 13.1
14:00 234.2 F 9.3
16:00 238.6 F 9.7
18:00 244.4 F 12.7
20:00 255.9 F 18.9
22:00 266.6 F 25.4
24:00 272.9 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 28.9
1:00 272.6 F 29.5
2:00 271.8 F 28.8
3:00 272.4 F 29.4
4:00 271.9 F 29.0

Min 271.7 F 28.7
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.6 F 0.0
5:00 230.2 F 0.0
10:00 233.7 F 0.0
15:00 232.0 F 0.0
20:00 228.3 F 0.0
25:00 233.2 F 0.0
30:00 229.5 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 24:14
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:02:21
CYCLE COMPLETE

06/06/2022 01:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____