

-----  
Midmark M11 - v1.0.0  
Total Cycles: 854

Name: RITTER 1  
Sterilizer ID: V2178198  
05/31/2021 01:35 PM

BEGIN PACKS CYCLE  
Temp: 250 Degrees F  
Time: 30 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 221.7 F 0.0  
2:00 216.2 F 5.1  
4:00 218.5 F 3.1  
6:00 228.1 F 5.4  
8:00 231.5 F 6.3  
10:00 237.4 F 8.8  
12:00 251.2 F 15.4

STERILIZING  
mm:ss Degrees PSI  
0:00 251.6 F 15.6  
1:00 252.2 F 15.8  
2:00 252.7 F 16.1  
3:00 251.8 F 15.6  
4:00 252.8 F 16.1  
5:00 252.0 F 15.7  
6:00 252.4 F 16.2  
7:00 252.6 F 16.0  
8:00 251.9 F 15.6  
9:00 252.5 F 16.2  
10:00 252.7 F 16.1  
11:00 252.2 F 15.8  
12:00 251.6 F 15.5  
13:00 252.5 F 16.2  
14:00 252.7 F 16.1  
15:00 252.3 F 15.9  
16:00 251.9 F 15.6  
17:00 251.8 F 15.8  
18:00 252.8 F 16.2  
19:00 252.6 F 16.0  
20:00 252.3 F 15.8  
21:00 251.9 F 15.6  
22:00 251.6 F 15.7  
23:00 252.6 F 16.2  
24:00 252.7 F 16.1  
25:00 252.4 F 16.0  
26:00 252.0 F 15.7  
27:00 251.7 F 15.5  
28:00 252.0 F 16.0  
29:00 252.8 F 16.2  
30:00 252.6 F 16.1

Min 251.5 F 15.5  
Max 252.8 F 16.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.2 F 0.0  
5:00 213.8 F 0.0  
10:00 214.8 F 0.0  
15:00 182.3 F 0.0  
20:00 183.2 F 0.0  
Dry Cycle Aborted

FILLING: 3:09  
HEATING: 12:36  
STERILIZING: 30:00  
VENTING: 2:56  
DRYING: 21:16  
TOTAL CYCLE: 01:09:57

05/31/2021 02:44 PM

ACCEPTED: X

REJECTED:

Operator: U01