
Midmark M11 - v1.0.0
Total Cycles: 4221

Name: RITTER 1
Sterilizer ID: V2178198
01/24/2024 08:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.2 F 0.0
2:00 97.7 F 0.2
4:00 109.7 F 1.1
6:00 124.7 F 3.0
8:00 146.7 F 5.6
10:00 167.6 F 8.9
12:00 209.0 F 12.3
14:00 211.8 F 3.7
16:00 218.5 F 7.1
18:00 226.8 F 10.6
20:00 231.1 F 7.8
22:00 235.0 F 9.6
24:00 239.9 F 10.2
26:00 243.0 F 11.8
28:00 254.4 F 17.5
30:00 264.1 F 23.4
32:00 271.9 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 271.8 F 28.4
2:00 272.5 F 28.7
3:00 272.3 F 29.0
4:00 271.4 F 28.7

Min 271.1 F 28.3
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.1 F 0.0
5:00 235.8 F 0.0
10:00 230.7 F 0.0
15:00 224.9 F 0.0
20:00 229.5 F 0.0
25:00 228.2 F 0.0
30:00 221.2 F 0.0

Drying Complete

FILLING: 0:56
HEATING: 32:08
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:10:23
CYCLE COMPLETE

01/24/2024 10:07 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____