
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001021
14:46 06/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:36
STERILIZING 3:46
136.3 C 223kPa 3:46
136.3 C 223kPa 4:16
136.4 C 225kPa 4:46
136.4 C 225kPa 5:16
136.4 C 225kPa 5:46
136.4 C 224kPa 6:16
136.3 C 224kPa 6:46
136.4 C 225kPa 7:16
136.4 C 225kPa 7:16
VENTING 7:16
AIR DRYING 8:05
CYCLE COMPLETE 20:30

Digital Signature #
6CD4E9CA566C639D
