
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001513
12:23 10/20/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:08
PRESSURIZING 6:49
STERILIZING 8:17
136.0 C 220kPa 8:17
135.9 C 218kPa 8:47
136.4 C 217kPa 9:17
136.2 C 221kPa 9:47
136.4 C 221kPa 10:17
136.5 C 221kPa 10:47
136.4 C 223kPa 11:17
136.0 C 219kPa 11:47
VENTING 11:47
AIR DRYING 12:49
CYCLE COMPLETE 28:33

Digital Signature #
7F7E277D72C35D69
