
Midmark M11 - v1.0.5
Total Cycles: 4422

Name: RITTER2
Sterilizer ID: V2313228
03/11/2025 05:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.5 F 0.0
2:00 170.8 F 3.2
4:00 196.7 F 9.0
6:00 216.5 F 6.5
8:00 232.4 F 10.9
10:00 233.7 F 8.7
12:00 237.9 F 9.5
14:00 249.9 F 15.7
16:00 263.3 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.7 F 29.5
2:00 271.9 F 29.2
3:00 272.7 F 29.4
4:00 271.8 F 29.0

Min 271.7 F 28.7
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.1 F 0.0
5:00 182.5 F 0.0
10:00 206.6 F 0.0
15:00 208.8 F 0.0
20:00 207.2 F 0.0
25:00 204.8 F 0.0
30:00 203.0 F 0.0

Drying Complete

FILLING: 2:14
HEATING: 17:57
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:24
CYCLE COMPLETE

03/11/2025 06:20 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____