
Midmark M11 - v1.0.5
Total Cycles: 4233

Name: RITTER2
Sterilizer ID: V2313228
12/20/2024 12:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.5 F 0.0
2:00 131.1 F 0.9
4:00 148.4 F 2.9
6:00 168.4 F 6.3
8:00 186.8 F 9.5
10:00 200.4 F 13.0
12:00 212.2 F 5.2
14:00 223.4 F 9.0
16:00 234.5 F 13.3
18:00 235.6 F 9.9
20:00 238.3 F 9.8
22:00 246.0 F 13.6
24:00 256.7 F 19.4
26:00 266.7 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.4 F 29.9
2:00 272.7 F 29.5
3:00 271.9 F 29.1
4:00 271.4 F 29.1

Min 271.4 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 202.4 F 0.0
10:00 202.8 F 0.0
15:00 193.4 F 0.0
20:00 187.5 F 0.0
25:00 181.9 F 0.0
30:00 188.3 F 0.0

Drying Complete

FILLING: 0:59
HEATING: 27:46
STERILIZING: 4:00
VENTING: 3:56
DRYING: 30:00
TOTAL CYCLE: 01:06:42
CYCLE COMPLETE

12/20/2024 01:13 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____