

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4276

Name: RITTER2  
Sterilizer ID: V2313228  
01/17/2025 02:29 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.8 F 0.0  
2:00 155.9 F 0.4  
4:00 182.8 F 2.8  
6:00 205.0 F 7.9  
8:00 221.9 F 13.5  
10:00 231.5 F 9.0  
12:00 236.6 F 10.0  
14:00 243.6 F 12.9  
16:00 242.8 F 12.0  
18:00 256.2 F 19.1  
20:00 268.1 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.8  
1:00 272.9 F 29.5  
2:00 271.7 F 28.8  
3:00 273.0 F 29.6  
4:00 272.1 F 28.8

Min 271.6 F 28.7  
Max 273.0 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.8 F 0.0  
5:00 211.6 F 0.0  
10:00 209.8 F 0.0  
15:00 209.7 F 0.0  
20:00 208.8 F 0.0  
25:00 210.1 F 0.0  
30:00 218.0 F 0.0  
Drying Complete

FILLING: 1:09  
HEATING: 21:14  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 00:59:42  
CYCLE COMPLETE

01/17/2025 03:28 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_