

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3508

Name: RITTER 1  
Sterilizer ID: V2178198  
07/10/2023 02:59 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 108.1 F 0.0  
2:00 111.8 F 0.0  
4:00 120.8 F 0.4  
6:00 149.6 F 2.6  
8:00 185.4 F 5.6  
10:00 205.4 F 9.8  
12:00 215.8 F 9.3  
14:00 225.9 F 7.8  
16:00 235.8 F 12.4  
18:00 235.4 F 9.2  
20:00 239.1 F 10.1  
22:00 245.9 F 13.3  
24:00 257.4 F 19.5  
26:00 268.0 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.2 F 28.7  
2:00 272.6 F 29.2  
3:00 271.4 F 28.5  
4:00 272.2 F 28.8

Min 271.4 F 28.4  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.9 F 0.0  
5:00 195.8 F 0.0  
10:00 209.3 F 0.0  
15:00 211.0 F 0.0  
20:00 209.1 F 0.0  
25:00 212.4 F 0.0  
30:00 214.5 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 27:18  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:05:34  
CYCLE COMPLETE

07/10/2023 04:04 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_