

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1841

Name: RITTER 1  
Sterilizer ID: V2178198  
04/19/2022 11:20 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.4 F 0.0  
2:00 195.9 F 4.4  
4:00 206.4 F 9.4  
6:00 219.4 F 6.6  
8:00 231.0 F 9.7  
10:00 237.5 F 8.2  
12:00 243.4 F 13.2  
14:00 242.9 F 12.0  
16:00 255.9 F 18.8  
18:00 267.5 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.7  
1:00 272.4 F 29.4  
2:00 271.6 F 28.7  
3:00 272.1 F 29.5  
4:00 272.4 F 29.2  
  
Min 271.2 F 28.6  
Max 272.5 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.5 F 0.0  
5:00 213.7 F 0.0  
10:00 221.7 F 0.0  
15:00 229.6 F 0.0  
20:00 231.4 F 0.0  
25:00 229.5 F 0.0  
30:00 229.2 F 0.0  
Drying Complete

FILLING: 1:35  
HEATING: 19:24  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:10  
CYCLE COMPLETE

04/19/2022 12:18 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_