

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5176

Name: RITTER 1  
Sterilizer ID: V2178198  
12/10/2024 06:42 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 93.6 F 0.0  
2:00 96.6 F 0.0  
4:00 124.4 F 0.0  
6:00 161.8 F 0.2  
8:00 190.5 F 0.4  
10:00 211.8 F 2.9  
12:00 221.7 F 6.5  
14:00 232.2 F 10.3  
16:00 233.4 F 8.3  
18:00 242.0 F 11.7  
20:00 238.7 F 9.6  
22:00 236.9 F 8.5  
24:00 250.8 F 15.5  
26:00 263.6 F 22.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.0  
1:00 272.8 F 28.7  
2:00 271.7 F 28.1  
3:00 272.8 F 28.7  
4:00 271.7 F 27.9

Min 271.6 F 27.9  
Max 273.0 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.1 F 0.0  
5:00 209.6 F 0.0  
10:00 193.9 F 0.0  
15:00 199.1 F 0.0  
20:00 196.4 F 0.0  
25:00 202.7 F 0.0  
30:00 204.0 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 27:59  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:06:01  
CYCLE COMPLETE

12/10/2024 07:47 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_