
Midmark M11 - v1.0.0
Total Cycles: 2806

Name: RITTER 1
Sterilizer ID: V2178198
01/23/2023 02:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.8 F 0.0
2:00 179.3 F 2.5
4:00 191.6 F 6.9
6:00 211.8 F 12.1
8:00 222.3 F 7.2
10:00 236.1 F 12.6
12:00 237.7 F 10.1
14:00 235.8 F 8.2
16:00 250.2 F 15.5
18:00 262.4 F 22.5
20:00 272.6 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.5
1:00 272.4 F 29.3
2:00 271.9 F 28.7
3:00 271.3 F 28.7
4:00 272.6 F 29.4

Min 271.1 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.8 F 0.0
5:00 213.6 F 0.0
10:00 232.0 F 0.0
15:00 232.2 F 0.0
20:00 227.5 F 0.0
25:00 224.9 F 0.0
Dry Cycle Aborted

FILLING: 1:17
HEATING: 20:18
STERILIZING: 4:00
VENTING: 3:19
DRYING: 27:56
TOTAL CYCLE: 00:56:49

01/23/2023 02:58 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____