
Midmark M11 - v1.0.5
Total Cycles: 1130

Name: RITTER2
Sterilizer ID: V2313228
05/27/2022 11:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.5 F 0.0
2:00 198.6 F 4.3
4:00 211.9 F 9.8
6:00 226.9 F 7.4
8:00 239.8 F 12.1
10:00 238.5 F 10.4
12:00 236.7 F 9.4
14:00 251.6 F 16.5
16:00 263.6 F 23.6
18:00 271.5 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.5 F 29.6
2:00 271.7 F 28.9
3:00 271.4 F 29.3
4:00 272.5 F 29.5

Min 271.0 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.2 F 0.0
5:00 228.0 F 0.0
10:00 241.8 F 0.0
15:00 234.6 F 0.0
20:00 225.5 F 0.0
25:00 224.3 F 0.0
30:00 224.3 F 0.0
Drying Complete

FILLING: 2:25
HEATING: 18:04
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:38
CYCLE COMPLETE

05/27/2022 12:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____