
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 008199
15:08 08/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:20
PRESSURIZING 2:23
STERILIZING 3:14
136.0 C 220kPa 3:14
136.1 C 220kPa 3:44
136.2 C 218kPa 4:14
136.4 C 223kPa 4:44
136.0 C 222kPa 5:14
136.3 C 220kPa 5:44
136.1 C 222kPa 6:14
136.4 C 225kPa 6:44
136.4 C 225kPa 6:44
VENTING 6:44
AIR DRYING 7:34
CYCLE COMPLETE 13:06

Digital Signature #
F5B5F6E39650810E
