
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001343
12:20 11/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:54
PRESSURIZING 3:01
STERILIZING 4:10
136.2 C 221kPa 4:10
136.4 C 225kPa 4:40
136.3 C 222kPa 5:10
136.2 C 223kPa 5:40
136.2 C 222kPa 6:10
136.3 C 224kPa 6:40
136.0 C 222kPa 7:10
136.4 C 224kPa 7:40
136.4 C 224kPa 7:40
VENTING 7:40
AIR DRYING 8:26
CYCLE COMPLETE 10:46

Digital Signature #
A411F39D613BFF44
