
Midmark M11 - v1.0.0
Total Cycles: 2358

Name: RITTER 1
Sterilizer ID: V2178198
09/22/2022 09:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.4 F 0.0
2:00 200.4 F 3.4
4:00 215.0 F 8.7
6:00 224.8 F 8.5
8:00 230.7 F 9.3
10:00 238.6 F 9.5
12:00 242.4 F 12.3
14:00 240.3 F 10.8
16:00 253.6 F 17.5
18:00 265.3 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.5 F 29.2
2:00 271.9 F 28.5
3:00 272.0 F 28.9
4:00 272.4 F 29.1

Min 271.3 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.9 F 0.0
5:00 230.7 F 0.0
10:00 203.2 F 0.0
15:00 234.5 F 0.0
20:00 228.6 F 0.0
25:00 220.4 F 0.0
30:00 209.8 F 0.0
Drying Complete

FILLING: 2:47
HEATING: 19:52
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:53
CYCLE COMPLETE

09/22/2022 10:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____