
Midmark M11 - v1.0.5
Total Cycles: 3125

Name: RITTER2
Sterilizer ID: V2313228
11/15/2023 10:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.0 F 0.0
2:00 72.8 F 0.0
4:00 95.9 F 0.4
6:00 108.8 F 1.7
8:00 141.3 F 4.6
10:00 168.5 F 8.3
12:00 187.2 F 12.2
14:00 208.5 F 4.6
16:00 218.9 F 8.5
18:00 229.3 F 12.5
20:00 231.5 F 8.5
22:00 241.5 F 12.1
24:00 242.6 F 12.1
26:00 254.7 F 18.4
28:00 265.5 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.7 F 29.7
2:00 271.3 F 28.9
3:00 272.5 F 29.8
4:00 271.9 F 29.1

Min 271.2 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.5 F 0.0
5:00 199.6 F 0.0
10:00 200.1 F 0.0
15:00 193.5 F 0.0
20:00 189.5 F 0.0
25:00 186.8 F 0.0
30:00 184.8 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 29:51
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:07:54
CYCLE COMPLETE

11/15/2023 11:27 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____