
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002380
15:49 13/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:24
STERILIZING 3:24
136.0 C 220kPa 3:24
136.4 C 221kPa 3:54
136.2 C 223kPa 4:24
136.0 C 223kPa 4:54
136.4 C 225kPa 5:24
136.5 C 224kPa 5:54
136.4 C 225kPa 6:24
136.2 C 222kPa 6:54
136.2 C 222kPa 6:54
VENTING 6:54
AIR DRYING 7:46
CYCLE COMPLETE 12:07

Digital Signature #
F39F68E4FAA94ABD
