
Midmark M11 - v1.0.5
Total Cycles: 2291

Name: RITTER2
Sterilizer ID: V2313228
04/24/2023 12:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.3 F 0.0
2:00 135.3 F 0.9
4:00 153.1 F 4.0
6:00 183.2 F 8.2
8:00 202.9 F 13.3
10:00 219.1 F 7.4
12:00 234.5 F 12.4
14:00 234.8 F 9.6
16:00 237.2 F 9.3
18:00 247.4 F 14.6
20:00 259.6 F 21.4
22:00 270.9 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.9 F 29.8
2:00 271.4 F 29.2
3:00 272.6 F 30.0
4:00 271.9 F 29.3

Min 271.4 F 29.1
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.0 F 0.0
5:00 204.6 F 0.0
10:00 226.4 F 0.0
15:00 220.8 F 0.0
20:00 221.5 F 0.0
25:00 232.5 F 0.0
30:00 229.3 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 22:45
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:56
CYCLE COMPLETE

04/24/2023 01:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____