
Midmark M11 - v1.0.5
Total Cycles: 1569

Name: RITTER2
Sterilizer ID: V2313228
10/05/2022 03:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.6 F 0.0
2:00 175.7 F 0.0
4:00 195.8 F 0.7
6:00 216.9 F 5.1
8:00 228.1 F 10.0
10:00 234.7 F 8.2
12:00 241.0 F 12.2
14:00 240.2 F 10.8
16:00 238.8 F 10.2
18:00 252.9 F 17.2
20:00 264.9 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.6 F 29.4
2:00 271.7 F 28.6
3:00 272.1 F 29.4
4:00 272.4 F 29.2

Min 271.4 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 216.9 F 0.0
10:00 229.6 F 0.0
15:00 234.4 F 0.0
20:00 228.3 F 0.0
25:00 226.8 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 21:51
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:05
CYCLE COMPLETE

10/05/2022 04:28 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____