
Midmark M11 - v1.0.0
Total Cycles: 5500

Name:
Sterilizer ID: V2178198
04/11/2025 01:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.5 F 0.0
2:00 157.7 F 2.8
4:00 167.9 F 6.4
6:00 185.8 F 10.3
8:00 222.0 F 5.6
10:00 214.8 F 6.2
12:00 226.6 F 10.2
14:00 235.3 F 8.7
16:00 241.4 F 11.5
18:00 238.2 F 9.4
20:00 250.5 F 15.4
22:00 261.0 F 21.3
24:00 270.7 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.8
1:00 272.0 F 28.8
2:00 272.3 F 29.0
3:00 272.3 F 29.2
4:00 272.1 F 29.3

Min 271.7 F 28.8
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 207.6 F 0.0
10:00 213.6 F 0.0
15:00 214.4 F 0.0
20:00 218.6 F 0.0
25:00 231.2 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 25:30
STERILIZING: 4:00
VENTING: 3:57
DRYING: 30:00
TOTAL CYCLE: 01:04:24
CYCLE COMPLETE

04/11/2025 02:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____