
Midmark M11 - v1.0.5
Total Cycles: 4324

Name: RITTER2
Sterilizer ID: V2313228
01/31/2025 01:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.0 F 0.0
2:00 157.4 F 1.8
4:00 156.7 F 4.7
6:00 170.7 F 8.4
8:00 192.1 F 12.5
10:00 215.2 F 5.6
12:00 224.8 F 9.6
14:00 232.7 F 10.1
16:00 236.7 F 11.0
18:00 235.2 F 9.0
20:00 246.6 F 15.0
22:00 257.9 F 20.9
24:00 267.2 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 30.4
1:00 272.1 F 30.6
2:00 272.2 F 30.8
3:00 272.1 F 30.8
4:00 272.1 F 30.8

Min 271.8 F 30.3
Max 272.7 F 30.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.8 F 0.0
5:00 182.7 F 0.0
10:00 201.5 F 0.0
15:00 200.5 F 0.0
20:00 194.0 F 0.0
25:00 189.5 F 0.0
30:00 186.9 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 25:24
STERILIZING: 4:00
VENTING: 3:59
DRYING: 30:00
TOTAL CYCLE: 01:04:23
CYCLE COMPLETE

01/31/2025 02:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____