

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 007682  
10:32 16/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:13  
PRESSURIZING 3:16  
STERILIZING 4:23  
136.4 C 223kPa 4:23  
136.3 C 222kPa 4:53  
136.3 C 222kPa 5:23  
136.4 C 224kPa 5:53  
136.2 C 223kPa 6:23  
136.4 C 220kPa 6:53  
136.2 C 219kPa 7:23  
136.5 C 223kPa 7:53  
136.5 C 223kPa 7:53  
VENTING 7:53  
AIR DRYING 8:47  
CYCLE COMPLETE 9:06

-----  
Digital Signature #  
CC967E0ECEFAE6C6  
-----