
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 000307
9:53 20/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 1:03
STOP BUTTON PRESSED
NOT STERILE
END TIME 1:03

Digital Signature #
4D2D208F05B22F76
