
Midmark M11 - v1.0.5
Total Cycles: 3042

Name: RITTER2
Sterilizer ID: V2313228
10/23/2023 11:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.4 F 0.0
2:00 158.3 F 2.7
4:00 173.6 F 6.7
6:00 197.4 F 11.5
8:00 217.1 F 5.4
10:00 226.2 F 10.7
12:00 230.5 F 7.8
14:00 239.5 F 11.2
16:00 246.2 F 13.7
18:00 258.8 F 20.7
20:00 270.4 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.4
1:00 272.5 F 29.5
2:00 272.8 F 29.9
3:00 271.7 F 29.4
4:00 272.2 F 29.6

Min 271.6 F 29.2
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.1 F 0.0
5:00 229.1 F 0.0
10:00 201.7 F 0.0
15:00 203.0 F 0.0
20:00 203.4 F 0.0
25:00 203.1 F 0.0
30:00 202.7 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 21:33
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:00:05
CYCLE COMPLETE

10/23/2023 12:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____