

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3694

Name: RITTER2  
Sterilizer ID: V2313228  
06/13/2024 10:22 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 135.5 F 0.0  
2:00 134.1 F 1.1  
4:00 136.6 F 3.2  
6:00 154.9 F 6.0  
8:00 177.3 F 9.6  
10:00 200.8 F 12.2  
12:00 215.3 F 7.4  
14:00 225.3 F 11.9  
16:00 230.1 F 8.0  
18:00 240.6 F 13.2  
20:00 240.6 F 11.4  
22:00 251.9 F 17.1  
24:00 261.8 F 23.1  
26:00 271.5 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.4  
1:00 272.0 F 30.3  
2:00 272.2 F 30.2  
3:00 272.0 F 30.2  
4:00 272.0 F 30.1

Min 270.8 F 29.2  
Max 273.0 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.4 F 0.0  
5:00 221.0 F 0.0  
10:00 234.4 F 0.0  
15:00 222.6 F 0.0  
20:00 210.1 F 0.0  
25:00 211.8 F 0.0  
30:00 214.9 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 26:33  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:05:02  
CYCLE COMPLETE

06/13/2024 11:27 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_