
Midmark M11 - v1.0.0
Total Cycles: 4710

Name: RITTER 1
Sterilizer ID: V2178198
07/10/2024 12:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.0 F 0.0
2:00 151.9 F 2.6
4:00 163.8 F 5.7
6:00 185.4 F 9.0
8:00 207.1 F 13.4
10:00 217.8 F 5.4
12:00 228.2 F 9.1
14:00 237.8 F 13.4
16:00 235.7 F 9.8
18:00 238.8 F 9.7
20:00 245.9 F 13.4
22:00 256.8 F 19.3
24:00 267.0 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.5 F 29.4
2:00 272.4 F 29.1
3:00 271.7 F 28.6
4:00 271.5 F 28.8

Min 271.3 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.6 F 0.0
5:00 195.7 F 0.0
10:00 193.9 F 0.0
15:00 190.9 F 0.0
20:00 188.8 F 0.0
25:00 193.1 F 0.0
30:00 196.7 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 25:39
STERILIZING: 4:00
VENTING: 4:02
DRYING: 30:00
TOTAL CYCLE: 01:04:34
CYCLE COMPLETE

07/10/2024 01:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____