
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000168
10:54 18/06/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:05
PRESSURIZING 3:08
STERILIZING 4:12
136.4 C 222kPa 4:12
136.4 C 225kPa 4:42
136.3 C 224kPa 5:12
136.2 C 223kPa 5:42
136.4 C 223kPa 6:12
136.1 C 221kPa 6:42
136.2 C 223kPa 7:12
136.2 C 223kPa 7:42
136.2 C 223kPa 7:42
VENTING 7:42
AIR DRYING 8:31
CYCLE COMPLETE 8:40

Digital Signature #
E13A669BDE2BF3A6
