
Midmark M11 - v1.0.0
Total Cycles: 2921

Name: RITTER 1
Sterilizer ID: V2178198
02/21/2023 09:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.3 F 0.0
2:00 182.9 F 3.5
4:00 191.8 F 7.8
6:00 210.0 F 12.7
8:00 221.3 F 7.4
10:00 234.7 F 12.8
12:00 235.9 F 9.7
14:00 237.6 F 8.6
16:00 248.2 F 14.7
18:00 260.4 F 21.5
20:00 271.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.5 F 29.4
2:00 271.6 F 28.6
3:00 272.3 F 29.5
4:00 272.2 F 29.1

Min 271.4 F 28.6
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 215.8 F 0.0
10:00 232.5 F 0.0
15:00 229.8 F 0.0
20:00 226.9 F 0.0
25:00 224.2 F 0.0
30:00 224.1 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 20:37
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:08
CYCLE COMPLETE

02/21/2023 10:46 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____