
Midmark M11 - v1.0.0
Total Cycles: 1359

Name: RITTER 1
Sterilizer ID: V2178198
11/11/2021 09:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.2 F 0.0
2:00 175.0 F 2.0
4:00 178.5 F 5.8
6:00 200.6 F 10.6
8:00 221.7 F 4.9
10:00 233.6 F 10.3
12:00 234.5 F 7.6
14:00 244.7 F 12.9
16:00 242.6 F 11.3
18:00 255.4 F 18.0
20:00 266.9 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.7 F 28.9
2:00 271.3 F 28.1
3:00 272.5 F 29.0
4:00 272.4 F 28.5

Min 271.1 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.4 F 0.0
5:00 182.5 F 0.0
10:00 180.3 F 0.0
15:00 175.1 F 0.0
20:00 171.9 F 0.0
25:00 168.0 F 0.0
30:00 167.1 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 21:30
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:00:00
CYCLE COMPLETE

11/11/2021 10:16 AM
ACCEPTED: X
REJECTED:
Operator: U02