
Midmark M11 - v1.0.0
Total Cycles: 3178

Name: RITTER 1
Sterilizer ID: V2178198
05/01/2023 07:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.7 F 0.0
2:00 77.7 F 0.0
4:00 102.4 F 0.0
6:00 110.8 F 0.0
8:00 144.7 F 1.7
10:00 169.9 F 3.6
12:00 199.5 F 6.3
14:00 210.7 F 9.9
16:00 220.4 F 11.4
18:00 224.9 F 7.6
20:00 235.4 F 11.7
22:00 232.8 F 8.3
24:00 243.1 F 12.4
26:00 243.0 F 12.2
28:00 254.7 F 18.3
30:00 265.3 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.5 F 29.6
2:00 271.6 F 28.8
3:00 272.4 F 29.7
4:00 272.1 F 29.2

Min 271.4 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.8 F 0.0
5:00 205.0 F 0.0
10:00 218.1 F 0.0
15:00 222.4 F 0.0
20:00 222.7 F 0.0
25:00 225.8 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:56
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:09:57
CYCLE COMPLETE

05/01/2023 08:41 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____