
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 000573
7:14 05/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:54
PRESSURIZING 3:55
STERILIZING 4:59
136.0 C 220kPa 4:59
136.4 C 224kPa 5:29
136.3 C 222kPa 5:59
136.3 C 223kPa 6:29
136.4 C 220kPa 6:59
136.4 C 222kPa 7:29
136.2 C 222kPa 7:59
136.4 C 223kPa 8:29
136.4 C 223kPa 8:29
VENTING 8:29
AIR DRYING 9:22
CYCLE COMPLETE 9:28

Digital Signature #
D73D8DED06E78DBD
