

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4865

Name: RITTER 1  
Sterilizer ID: V2178198  
09/04/2024 08:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.2 F 0.0  
2:00 70.0 F 0.0  
4:00 77.1 F 0.0  
6:00 98.2 F 0.0  
8:00 141.5 F 0.2  
10:00 167.2 F 0.6  
12:00 199.2 F 1.4  
14:00 209.2 F 4.6  
16:00 218.9 F 7.4  
18:00 228.5 F 10.6  
20:00 232.8 F 8.3  
22:00 238.0 F 10.4  
24:00 235.6 F 8.1  
26:00 246.7 F 13.4  
28:00 243.8 F 11.8  
30:00 255.9 F 18.0  
32:00 266.4 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 27.8  
1:00 272.5 F 28.5  
2:00 271.4 F 27.7  
3:00 272.5 F 28.5  
4:00 271.6 F 27.8  
  
Min 271.3 F 27.7  
Max 272.6 F 28.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.0 F 0.0  
5:00 225.6 F 0.0  
10:00 235.3 F 0.0  
15:00 228.1 F 0.0  
20:00 227.1 F 0.0  
25:00 228.4 F 0.0  
30:00 227.5 F 0.0  
  
Drying Complete

FILLING: 0:43  
HEATING: 33:40  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:11:50  
CYCLE COMPLETE

09/04/2024 09:34 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_