
Midmark M11 - v1.0.5
Total Cycles: 1232

Name: RITTER2
Sterilizer ID: V2313228
06/28/2022 08:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.2 F 0.0
2:00 153.3 F 1.5
4:00 168.1 F 4.9
6:00 190.3 F 9.7
8:00 219.6 F 6.0
10:00 224.8 F 8.0
12:00 237.3 F 13.0
14:00 236.6 F 10.2
16:00 236.8 F 8.6
18:00 248.7 F 15.1
20:00 260.3 F 21.6
22:00 270.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.6 F 29.7
2:00 271.4 F 28.9
3:00 272.0 F 29.8
4:00 273.4 F 30.2

Min 270.9 F 28.9
Max 273.4 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.1 F 0.0
5:00 221.1 F 0.0
10:00 222.4 F 0.0
15:00 226.4 F 0.0
20:00 193.2 F 0.0
25:00 188.9 F 0.0
30:00 185.8 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 22:47
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:01:33
CYCLE COMPLETE

06/28/2022 09:04 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____