
Midmark M11 - v1.0.5
Total Cycles: 2436

Name: RITTER2
Sterilizer ID: V2313228
05/30/2023 01:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.1 F 0.0
2:00 179.5 F 3.5
4:00 185.1 F 7.8
6:00 209.5 F 13.5
8:00 225.9 F 9.1
10:00 233.9 F 9.3
12:00 240.6 F 12.6
14:00 240.8 F 11.4
16:00 254.3 F 18.3
18:00 266.1 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.7 F 29.8
2:00 271.4 F 29.4
3:00 272.7 F 29.9
4:00 271.7 F 29.3

Min 271.3 F 29.2
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.4 F 0.0
5:00 214.0 F 0.0
10:00 226.8 F 0.0
15:00 230.0 F 0.0
20:00 228.5 F 0.0
25:00 229.5 F 0.0
30:00 228.3 F 0.0

Drying Complete

FILLING: 1:35
HEATING: 19:37
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:24
CYCLE COMPLETE

05/30/2023 01:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____