

-----  
Midmark M11 - v1.0.5  
Total Cycles: 756

Name: RITTER2  
Sterilizer ID: V2313228  
02/03/2022 07:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.9 F 0.0  
2:00 75.2 F 0.0  
4:00 79.4 F 0.0  
6:00 100.9 F 0.1  
8:00 129.9 F 0.3  
10:00 165.3 F 0.9  
12:00 187.3 F 4.1  
14:00 202.3 F 7.5  
16:00 217.5 F 11.4  
18:00 226.9 F 6.7  
20:00 237.6 F 11.7  
22:00 235.7 F 9.2  
24:00 239.2 F 9.9  
26:00 247.6 F 14.3  
28:00 259.0 F 20.7  
30:00 269.9 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.6  
1:00 272.4 F 29.2  
2:00 271.2 F 28.5  
3:00 272.4 F 29.3  
4:00 271.5 F 28.5

Min 271.1 F 28.4  
Max 272.5 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 175.6 F 0.0  
5:00 226.5 F 0.0  
10:00 203.8 F 0.0  
15:00 185.2 F 0.0  
20:00 193.4 F 0.0  
25:00 216.8 F 0.0  
30:00 222.9 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 30:53  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 01:08:59  
CYCLE COMPLETE

02/03/2022 08:51 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_