
Midmark M11 - v1.0.0
Total Cycles: 2590

Name: RITTER 1
Sterilizer ID: V2178198
11/21/2022 04:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 222.8 F 0.0
2:00 221.3 F 6.5
4:00 228.9 F 13.2
6:00 234.4 F 10.0
8:00 236.5 F 7.8
10:00 240.4 F 10.0
12:00 249.5 F 14.9
14:00 262.4 F 22.2
16:00 272.0 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.5 F 28.8
2:00 271.5 F 28.1
3:00 272.6 F 29.0
4:00 272.1 F 28.5

Min 271.4 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 242.5 F 0.0
10:00 233.5 F 0.0
15:00 229.7 F 0.0
20:00 227.6 F 0.0
25:00 230.4 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 2:38
HEATING: 16:09
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:56:02
CYCLE COMPLETE

11/21/2022 05:51 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____