
Midmark M11 - v1.0.0
Total Cycles: 853

Name: RITTER 1
Sterilizer ID: V2178198
05/31/2021 12:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.1 F 0.0
2:00 191.6 F 3.5
4:00 209.8 F 9.1
6:00 222.0 F 7.2
8:00 235.2 F 10.4
10:00 234.0 F 7.9
12:00 241.9 F 10.7
14:00 248.7 F 14.1
16:00 261.8 F 21.5
18:00 272.5 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.8
1:00 272.3 F 28.1
2:00 272.4 F 28.5
3:00 271.9 F 27.9
4:00 272.7 F 28.6

Min 271.4 F 27.6
Max 272.8 F 28.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.3 F 0.0
5:00 203.2 F 0.0
10:00 202.6 F 0.0
Dry Cycle Aborted

FILLING: 2:21
HEATING: 18:16
STERILIZING: 4:00
VENTING: 3:07
DRYING: 14:58
TOTAL CYCLE: 00:42:42

05/31/2021 01:33 PM
ACCEPTED: X
REJECTED:
Operator: U01