
Midmark M11 - v1.0.5
Total Cycles: 2079

Name: RITTER2
Sterilizer ID: V2313228
02/22/2023 03:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.9 F 0.0
2:00 192.0 F 3.7
4:00 209.1 F 9.2
6:00 224.8 F 8.2
8:00 235.8 F 11.3
10:00 236.2 F 9.5
12:00 238.2 F 9.1
14:00 250.9 F 16.0
16:00 263.5 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.3 F 29.2
2:00 271.2 F 29.0
3:00 272.2 F 29.2
4:00 271.1 F 28.6

Min 271.0 F 28.6
Max 272.4 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.6 F 0.0
5:00 221.9 F 0.0
10:00 234.3 F 0.0
15:00 228.3 F 0.0
20:00 228.1 F 0.0
25:00 228.2 F 0.0
30:00 226.8 F 0.0
Drying Complete

FILLING: 2:06
HEATING: 17:58
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:18
CYCLE COMPLETE

02/22/2023 04:13 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____