

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2373

Name: RITTER2  
Sterilizer ID: V2313228  
05/16/2023 09:51 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 201.9 F 0.0  
2:00 194.9 F 3.3  
4:00 199.4 F 8.4  
6:00 221.8 F 10.0  
8:00 231.0 F 10.3  
10:00 231.9 F 7.9  
12:00 240.4 F 11.1  
14:00 246.3 F 14.1  
16:00 259.1 F 21.1  
18:00 270.4 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.3  
1:00 272.7 F 29.9  
2:00 271.5 F 29.5  
3:00 272.2 F 29.6  
4:00 271.7 F 29.7  
  
Min 271.3 F 29.1  
Max 272.7 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.7 F 0.0  
5:00 217.2 F 0.0  
10:00 230.1 F 0.0  
15:00 227.2 F 0.0  
20:00 224.9 F 0.0  
25:00 222.5 F 0.0  
30:00 221.1 F 0.0  
Drying Complete

FILLING: 2:10  
HEATING: 18:41  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 00:58:08  
CYCLE COMPLETE

05/16/2023 10:49 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_