

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002390  
14:29 16/12/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:22  
PRESSURIZING 2:26  
STERILIZING 3:35  
136.4 C 224kPa 3:35  
136.4 C 225kPa 4:05  
136.3 C 224kPa 4:35  
136.4 C 225kPa 5:05  
136.3 C 224kPa 5:35  
136.4 C 224kPa 6:05  
136.3 C 225kPa 6:35  
136.4 C 225kPa 7:05  
136.4 C 225kPa 7:05  
VENTING 7:05  
AIR DRYING 7:58  
CYCLE COMPLETE 15:01  
-----

Digital Signature #  
5DA379FFF5A89387  
-----