
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007637
7:14 05/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:55
PRESSURIZING 2:58
STERILIZING 4:00
136.4 C 220kPa 4:00
136.4 C 223kPa 4:30
136.4 C 220kPa 5:00
136.2 C 219kPa 5:30
136.4 C 221kPa 6:00
136.3 C 221kPa 6:30
136.3 C 220kPa 7:00
135.9 C 218kPa 7:30
135.9 C 218kPa 7:30
VENTING 7:30
AIR DRYING 8:19
CYCLE COMPLETE 38:19

Digital Signature #
7813E4770FC720C2
