
Midmark M11 - v1.0.5
Total Cycles: 348

Name: RITTER2
Sterilizer ID: V2313228
09/09/2021 01:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.8 F 0.0
2:00 162.0 F 2.9
4:00 181.9 F 7.1
6:00 201.4 F 12.7
8:00 221.2 F 7.6
10:00 235.0 F 13.1
12:00 237.5 F 11.0
14:00 237.7 F 10.0
16:00 251.8 F 17.0
18:00 263.8 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 272.0 F 29.3
2:00 272.4 F 29.6
3:00 272.6 F 30.0
4:00 271.7 F 29.5

Min 271.6 F 29.2
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 206.8 F 0.0
10:00 218.6 F 0.0
15:00 217.6 F 0.0
20:00 215.1 F 0.0
25:00 211.7 F 0.0
30:00 204.3 F 0.0

Drying Complete

FILLING: 2:46
HEATING: 19:55
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:47
CYCLE COMPLETE

09/09/2021 02:27 PM
ACCEPTED: X
REJECTED:
Operator: U02