
Midmark M11 - v1.0.0
Total Cycles: 3487

Name: RITTER 1
Sterilizer ID: V2178198
07/05/2023 10:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.6 F 0.0
2:00 187.3 F 3.0
4:00 192.7 F 7.0
6:00 212.5 F 12.0
8:00 223.2 F 6.8
10:00 234.0 F 11.8
12:00 234.7 F 8.9
14:00 240.1 F 10.3
16:00 246.3 F 13.5
18:00 258.3 F 19.9
20:00 269.0 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.4 F 29.1
2:00 272.1 F 28.6
3:00 271.3 F 28.3
4:00 272.3 F 29.2

Min 271.2 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.2 F 0.0
5:00 225.1 F 0.0
10:00 229.5 F 0.0
15:00 231.0 F 0.0
20:00 229.0 F 0.0
25:00 229.3 F 0.0
30:00 225.9 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 21:11
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:00:16
CYCLE COMPLETE

07/05/2023 11:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____