
Midmark M11 - v1.0.0
Total Cycles: 4262

Name: RITTER 1
Sterilizer ID: V2178198
02/02/2024 10:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.7 F 0.0
2:00 145.2 F 1.8
4:00 156.6 F 4.7
6:00 181.8 F 8.6
8:00 202.4 F 13.3
10:00 216.6 F 5.8
12:00 227.4 F 10.1
14:00 234.0 F 8.3
16:00 241.5 F 11.8
18:00 239.4 F 9.9
20:00 252.3 F 16.3
22:00 263.5 F 22.8
24:00 272.0 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.1
1:00 272.7 F 29.3
2:00 272.1 F 28.5
3:00 272.4 F 29.3
4:00 272.1 F 28.8

Min 271.0 F 28.0
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.9 F 0.0
5:00 214.8 F 0.0
10:00 226.4 F 0.0
15:00 233.1 F 0.0
20:00 209.1 F 0.0
25:00 220.6 F 0.0
30:00 218.9 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 24:11
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:02:40
CYCLE COMPLETE

02/02/2024 11:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____